

6003 Chapel Hill Road · Suite # 153 · Raleigh, NC 27607

## **Certificate of Analysis**

Product:	Goat anti-Cystatin C, Ig-Fraction	
Part Number: Lot Number: Concentration: Amount: Form: Purification: Purification: Purity: Host: Immunogen: Buffer: Preservative: Storage: Expiration Date:	GtxHu-042-B 31-26-090512 22.5 mg/ml (E <sup>1%</sup> at 280 nm = 13.0) 1.0 ml Clear, colorless liquid, 0.2 $\mu$ m filtered Purified using a proprietary fractionation process $\geq 80 \%$ based on SDS-PAGE Goat Purified Human Cystatin C from urine 10 mM Sodium Phosphate, 0.15M Sodium Chloride, pH 7.2 0.05 % (w/v) Sodium Azide 2-8 °C October 22, 2021	
Specificity:	Human Cystatin C	
Country of Origin:	Goat serum was obtained from healthy animals of US origin and under the care of a registered veterinarian.	
Applications:	This antibody is suitable for use in most immunoassay formats for the detection of Human Cystatin C. The optimal working dilution should be determined by the investigator.	
Disclaimer:	For in vitro Laboratory Use Only. Not for diagnostic or therapeutic use. Not for human or animal consumption. The proper selection and use of our products is the sole responsibility of the end user, and therefore we can offer no guarantee to a specific experimental outcome. Suggested applications of our products are not recommendations to use our products in violation of any patent or as a license under any patent of ImmunoReagents, Inc. <b>Product may not be resold or modified for resale without prior written</b>	
approval of ImmunoReagents, Inc.		

## \*ImmunoReagents,Inc.

6003 Chapel Hill Road · Suite # 153 · Raleigh, NC 27607

## **Testing Results**

PRODUCT: PART NUMBER: LOT NUMBER:	Goat anti-Cystatin C, Ig-Fraction GtxHu-042-B 31-26-090512		
<u>TESTS</u>	SPECIFICATIONS	<b>RESULTS</b>	
Appearance:	Clear, colorless liquid	PASS	
Purity:	Two predominant bands at 50 and 24 kDa based on SDS-PAGE	PASS	
SDS-PAGE			
1 2			
<b>5</b> 5k	Da		
<b>3</b> 1k	Da		
21kDa			
-			
	2-B, Lot # 31-26-090512 r Weight Standard		
Date: October 23, 2	019 Signature:	and	

Quality Control